



**American College of Radiology
Accreditation Phantom for Whole Body Scanners**

Manufactured by 

11525 Sorrento Valley Road
Suite - B
San Diego, CA 92121
USA
T: (858) 794-7200
F: (858) 704-4959
E: customerservice @ jmspecialtyparts.com

Item: ACR-PH1 (Large Phantom)

Specifications as designated by the American College of Radiology



The phantom is an MRI accessory used to evaluate the performance of an MRI machine.
HTS: 9018.19.4000

The ACR-MRI phantom is an 8" (203.2mm) diameter cylinder, 6.82" (173.40mm) tall with various fixtures inside including a dinosaur, drilled plate grid, membrane stack, ramp, resolution block, and S-N chamber.

The level bar is 7.00" (177.80mm) length, .375" (9.53mm) width, .375" (9.53mm) height.

Inside the resolution block is ~1 ml hydrogenated vegetable oil. (brand name – Crisco)

The plate has nine holes, each hole is 1.5mm. Holes are spaced 60mm center to center.

The measured volume of one, version 2 phantom is 3627 ml.
The volumes could be +/- 10 ml, maybe more, but likely less.

The salt solution inside the phantom contains <11 grams nickel chloride hexahydrate and 17 grams of sodium chloride in 3.8 liters of water.

This unit may break open if subjected to extremes of temperature.
Keep temperature between 60-80°F (15-27°C).

MRI PHANTOMS CONTAIN NONE OF THE FOLLOWING:

- Lead (Pb)
- Mercury (Hg)
- Cadmium (Cd)
- Hexavalent Chromium (Cr VI)
- Polybrominated Biphenyls (PBB)
- Polybrominated Diphenyl Ethers (PBDE)

FDA has no code for MRI accessories.

These devices are not medical devices according to the European medical device regulation 2017/745. Therefore it is not allowed to label these products with a CE mark.